
Midmark M11 - v1.0.0
Total Cycles: 3218

Name: RITTER 1
Sterilizer ID: V2178198
05/11/2023 09:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.7 F 0.0
2:00 170.8 F 0.0
4:00 188.8 F 3.0
6:00 208.2 F 7.7
8:00 221.8 F 12.4
10:00 228.1 F 7.8
12:00 240.0 F 12.9
14:00 239.2 F 10.4
16:00 235.9 F 8.5
18:00 250.5 F 15.3
20:00 262.1 F 21.9
22:00 272.4 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.4 F 28.6
2:00 271.8 F 28.5
3:00 272.2 F 28.5
4:00 271.8 F 28.5

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.1 F 0.0
5:00 207.7 F 0.0
10:00 216.4 F 0.0
15:00 212.7 F 0.0
20:00 209.3 F 0.0
25:00 206.4 F 0.0
30:00 206.0 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 22:25
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:01:08
CYCLE COMPLETE

05/11/2023 11:00 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____