
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000374
14:25 31/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:28
PRESSURIZING 3:29
STERILIZING 4:31
136.0 C 220kPa 4:31
136.1 C 222kPa 5:01
136.2 C 220kPa 5:31
136.4 C 225kPa 6:01
136.4 C 224kPa 6:31
136.1 C 223kPa 7:01
136.3 C 224kPa 7:31
136.1 C 221kPa 8:01
136.1 C 221kPa 8:01
VENTING 8:02
AIR DRYING 8:55
CYCLE COMPLETE 9:30

Digital Signature #
991FBA75E07AE733
