

-----  
Midmark M11 - v1.0.0  
Total Cycles: 988

Name: RITTER 1  
Sterilizer ID: V2178198  
07/08/2021 03:22 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 195.4 F 0.0  
2:00 195.7 F 6.2  
4:00 232.4 F 12.6  
6:00 227.4 F 8.5  
8:00 236.6 F 9.0  
10:00 245.1 F 12.9  
12:00 243.2 F 11.7  
14:00 257.7 F 19.2  
16:00 270.1 F 27.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.0  
1:00 271.8 F 28.1  
2:00 272.4 F 28.5  
3:00 272.2 F 28.8  
4:00 271.2 F 27.9  
  
Min 271.2 F 27.9  
Max 272.5 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.1 F 0.0  
5:00 238.8 F 0.0  
10:00 240.3 F 0.0  
15:00 231.4 F 0.0  
20:00 226.9 F 0.0  
25:00 227.8 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 0:36  
HEATING: 16:46  
STERILIZING: 4:00  
VENTING: 2:50  
DRYING: 30:00  
TOTAL CYCLE: 00:54:12  
CYCLE COMPLETE

07/08/2021 04:16 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02