
Midmark M11 - v1.0.0
Total Cycles: 1583

Name: RITTER 1
Sterilizer ID: V2178198
02/01/2022 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.1 F 0.0
2:00 176.4 F 4.0
4:00 192.9 F 8.8
6:00 214.6 F 7.7
8:00 225.7 F 8.5
10:00 236.4 F 11.0
12:00 240.6 F 11.4
14:00 239.0 F 9.8
16:00 252.8 F 16.6
18:00 264.6 F 23.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.8 F 29.0
2:00 272.1 F 28.4
3:00 271.4 F 28.2
4:00 272.7 F 29.1

Min 271.4 F 28.0
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.0 F 0.0
5:00 231.9 F 0.0
10:00 225.8 F 0.0
15:00 225.2 F 0.0
20:00 224.8 F 0.0
25:00 226.0 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:33
HEATING: 19:55
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:58:45
CYCLE COMPLETE

02/01/2022 03:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____