
Midmark M11 - v1.0.0
Total Cycles: 2481

Name: RITTER 1
Sterilizer ID: V2178198
10/24/2022 04:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.7 F 0.0
2:00 200.4 F 4.2
4:00 209.6 F 9.4
6:00 221.1 F 6.2
8:00 230.0 F 9.5
10:00 238.9 F 9.1
12:00 242.5 F 12.4
14:00 241.4 F 11.0
16:00 254.9 F 17.9
18:00 266.7 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.0
2:00 271.8 F 28.3
3:00 271.4 F 28.4
4:00 272.4 F 29.0

Min 271.2 F 28.0
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 226.9 F 0.0
10:00 236.2 F 0.0
15:00 232.0 F 0.0
20:00 225.4 F 0.0
25:00 223.4 F 0.0
30:00 226.0 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 19:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:42
CYCLE COMPLETE

10/24/2022 05:15 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____