
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001221
15:25 18/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:48
PRESSURIZING 2:57
STERILIZING 4:12
136.0 C 220kPa 4:12
135.9 C 221kPa 4:42
136.2 C 223kPa 5:12
136.4 C 222kPa 5:42
136.4 C 224kPa 6:12
136.1 C 222kPa 6:42
136.5 C 226kPa 7:12
136.4 C 225kPa 7:42
136.4 C 225kPa 7:42
VENTING 7:42
AIR DRYING 8:33
CYCLE COMPLETE 23:33

Digital Signature #
37351DA2CFD1C7FA
