

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 001414  
9:29 09/09/2021  
UNWRAPPED  
134 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:58  
PRESSURIZING 5:24  
STERILIZING 6:47  
136.0 C 220kPa 6:47  
136.1 C 220kPa 7:17  
136.5 C 223kPa 7:47  
136.5 C 221kPa 8:17  
136.2 C 221kPa 8:47  
136.1 C 220kPa 9:18  
136.4 C 222kPa 9:48  
136.1 C 220kPa 10:17  
VENTING 10:17  
AIR DRYING 11:17  
CYCLE COMPLETE 11:26

-----  
Digital Signature #  
3F2A44A3FF75D67E  
-----