
Midmark M11 - v1.0.5
Total Cycles: 4311

Name: RITTER2
Sterilizer ID: V2313228
01/29/2025 01:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 160.2 F 0.0
2:00 161.5 F 2.4
4:00 173.6 F 6.1
6:00 196.5 F 10.8
8:00 223.2 F 5.2
10:00 226.5 F 10.8
12:00 232.0 F 8.0
14:00 241.2 F 11.7
16:00 244.2 F 13.1
18:00 256.5 F 19.5
20:00 267.4 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 30.1
1:00 271.6 F 30.3
2:00 271.8 F 30.5
3:00 271.7 F 30.5
4:00 271.6 F 30.4

Min 271.3 F 30.0
Max 272.7 F 30.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 180.6 F 0.0
5:00 182.7 F 0.0
10:00 180.9 F 0.0
15:00 176.9 F 0.0
20:00 174.2 F 0.0
25:00 175.7 F 0.0
30:00 164.9 F 0.0
Drying Complete

FILLING: 1:15
HEATING: 22:00
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:00:57
CYCLE COMPLETE

01/29/2025 02:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____