

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4985

Name: RITTER 1  
Sterilizer ID: V2178198  
10/10/2024 10:59 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 164.0 F 0.0  
2:00 174.4 F 0.6  
4:00 188.9 F 2.7  
6:00 212.0 F 6.8  
8:00 224.9 F 10.9  
10:00 225.3 F 5.7  
12:00 234.9 F 9.7  
14:00 238.9 F 9.7  
16:00 243.2 F 11.9  
18:00 240.0 F 10.3  
20:00 252.6 F 16.3  
22:00 263.5 F 22.7  
24:00 272.4 F 28.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.1  
1:00 272.6 F 28.9  
2:00 272.2 F 28.5  
3:00 271.2 F 28.0  
4:00 271.8 F 28.7

Min 271.1 F 28.0  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.6 F 0.0  
5:00 234.5 F 0.0  
10:00 258.8 F 0.0  
15:00 247.5 F 0.0  
20:00 216.6 F 0.0  
25:00 215.7 F 0.0  
30:00 214.6 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 24:19  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:03:08  
CYCLE COMPLETE

10/10/2024 12:02 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_