

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3874

Name: RITTER2  
Sterilizer ID: V2313228  
08/19/2024 03:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.2 F 0.0  
2:00 147.1 F 2.4  
4:00 172.9 F 6.3  
6:00 197.9 F 11.7  
8:00 222.1 F 6.3  
10:00 233.5 F 12.2  
12:00 235.9 F 9.8  
14:00 235.4 F 8.6  
16:00 250.0 F 15.7  
18:00 262.3 F 22.9  
20:00 271.7 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 273.1 F 30.1  
2:00 271.9 F 29.4  
3:00 272.1 F 29.9  
4:00 273.3 F 30.0

Min 271.4 F 29.2  
Max 273.3 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.6 F 0.0  
5:00 217.5 F 0.0  
10:00 191.7 F 0.0  
15:00 187.0 F 0.0  
20:00 183.0 F 0.0  
25:00 180.2 F 0.0  
30:00 177.5 F 0.0  
Drying Complete

FILLING: 1:18  
HEATING: 20:55  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 00:59:32  
CYCLE COMPLETE

08/19/2024 04:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_