
Midmark M11 - v1.0.5
Total Cycles: 2162

Name: RITTER2
Sterilizer ID: V2313228
03/20/2023 11:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.9 F 0.0
2:00 172.2 F 3.7
4:00 192.6 F 8.2
6:00 214.1 F 10.0
8:00 227.8 F 9.6
10:00 233.3 F 8.7
12:00 243.8 F 13.0
14:00 244.5 F 13.0
16:00 257.7 F 20.1
18:00 269.6 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 271.7 F 29.6
2:00 272.8 F 29.8
3:00 271.9 F 29.3
4:00 271.1 F 29.0

Min 271.0 F 28.9
Max 272.9 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 196.4 F 0.0
5:00 229.5 F 0.0
10:00 227.9 F 0.0
15:00 229.6 F 0.0
20:00 224.5 F 0.0
25:00 222.1 F 0.0
30:00 223.2 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 18:55
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:57:48
CYCLE COMPLETE

03/20/2023 12:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____