
Midmark M11 - v1.0.0
Total Cycles: 926

Name: RITTER 1
Sterilizer ID: V2178198
06/22/2021 02:30 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.5 F 0.0
2:00 109.1 F 0.0
4:00 126.0 F 0.1
6:00 162.6 F 0.7
8:00 185.4 F 3.5
10:00 202.0 F 6.7
12:00 223.2 F 10.2
14:00 229.7 F 7.4
16:00 234.0 F 10.0
18:00 234.4 F 9.0
20:00 243.3 F 12.7
22:00 241.7 F 11.3
24:00 254.2 F 17.6
26:00 265.2 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.7 F 29.0
2:00 271.2 F 28.3
3:00 272.5 F 29.2
4:00 271.9 F 28.6

Min 271.1 F 28.3
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.8 F 0.0
5:00 227.9 F 0.0
10:00 209.4 F 0.0
15:00 199.3 F 0.0
20:00 225.2 F 0.0
25:00 233.6 F 0.0
30:00 228.6 F 0.0

Drying Complete

FILLING: 1:04
HEATING: 27:53
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 01:06:17
CYCLE COMPLETE

06/22/2021 03:36 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____