

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3560

Name: RITTER2  
Sterilizer ID: V2313228  
04/25/2024 12:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 91.4 F 0.0  
2:00 92.1 F 0.5  
4:00 102.7 F 2.2  
6:00 124.9 F 4.5  
8:00 151.0 F 7.6  
10:00 172.7 F 11.1  
12:00 208.4 F 5.1  
14:00 216.0 F 7.2  
16:00 228.1 F 11.2  
18:00 228.1 F 6.7  
20:00 238.4 F 11.9  
22:00 236.9 F 9.8  
24:00 248.7 F 15.4  
26:00 259.4 F 21.3  
28:00 269.0 F 27.5  
30:00 271.8 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.9  
1:00 272.9 F 30.2  
2:00 272.1 F 30.3  
3:00 272.4 F 29.9  
4:00 272.3 F 30.5

Min 271.7 F 29.7  
Max 273.0 F 30.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 198.3 F 0.0  
5:00 213.9 F 0.0  
10:00 232.0 F 0.0  
15:00 230.0 F 0.0  
20:00 220.9 F 0.0  
25:00 217.0 F 0.0  
30:00 210.4 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 30:02  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:08:35  
CYCLE COMPLETE

04/25/2024 01:57 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_