
Midmark M11 - v1.0.5
Total Cycles: 171

Name: RITTER2
Sterilizer ID: V2313228
07/07/2021 08:07 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.0 F 0.0
2:00 153.6 F 1.9
4:00 161.0 F 5.4
6:00 187.6 F 9.9
8:00 208.2 F 5.3
10:00 223.3 F 8.1
12:00 235.0 F 13.2
14:00 238.3 F 10.9
16:00 237.2 F 9.6
18:00 250.9 F 16.2
20:00 262.4 F 22.8
22:00 272.6 F 29.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.0
1:00 272.5 F 29.5
2:00 271.8 F 29.3
3:00 272.4 F 29.4
4:00 271.7 F 29.2

Min 271.5 F 28.9
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.3 F 0.0
5:00 205.8 F 0.0
10:00 199.4 F 0.0
15:00 198.1 F 0.0
20:00 200.3 F 0.0
25:00 214.2 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 1:05
HEATING: 22:24
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:39
CYCLE COMPLETE

07/07/2021 09:07 AM
ACCEPTED: X
REJECTED:
Operator: U02