
Midmark M11 - v1.0.0
Total Cycles: 3679

Name: RITTER 1
Sterilizer ID: V2178198
08/21/2023 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 67.8 F 0.0
2:00 79.4 F 0.0
4:00 77.3 F 0.0
6:00 106.8 F 1.6
8:00 142.0 F 4.2
10:00 168.1 F 7.2
12:00 186.3 F 11.1
14:00 203.1 F 2.9
16:00 219.8 F 6.1
18:00 230.1 F 9.7
20:00 237.7 F 11.6
22:00 237.8 F 9.7
24:00 238.1 F 9.3
26:00 246.1 F 13.1
28:00 257.9 F 19.4
30:00 268.4 F 25.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.1
1:00 271.6 F 27.9
2:00 271.6 F 28.0
3:00 271.8 F 28.2
4:00 272.2 F 28.4

Min 271.4 F 27.9
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.6 F 0.0
5:00 207.9 F 0.0
10:00 208.1 F 0.0
15:00 223.5 F 0.0
20:00 234.1 F 0.0
25:00 224.5 F 0.0
30:00 225.5 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 31:09
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:09:08
CYCLE COMPLETE

08/21/2023 12:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____