

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001055  
13:08 04/05/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:16  
PRESSURIZING 3:29  
STERILIZING 5:06  
136.0 C 220kPa 5:06  
136.4 C 223kPa 5:36  
136.3 C 223kPa 6:06  
136.4 C 225kPa 6:36  
136.3 C 225kPa 7:06  
136.2 C 223kPa 7:36  
136.3 C 223kPa 8:06  
136.4 C 225kPa 8:36  
136.4 C 225kPa 8:36  
VENTING 8:36  
CYCLE ABORTED 8:47  
STOP BUTTON PRESSED  
NOT STERILE  
END TIME 8:47  
-----

Digital Signature #  
DE56E69EEC905E87  
-----