

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002606  
11:35 25/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:13  
PRESSURIZING 3:14  
STERILIZING 4:25  
136.5 C 226kPa 4:25  
136.3 C 225kPa 4:55  
136.4 C 222kPa 5:25  
136.4 C 225kPa 5:55  
136.1 C 223kPa 6:25  
136.1 C 223kPa 6:55  
136.4 C 225kPa 7:25  
136.4 C 226kPa 7:55  
136.4 C 226kPa 7:55  
VENTING 7:55  
AIR DRYING 8:47  
CYCLE COMPLETE 9:25  
-----

Digital Signature #  
E16BF68EF8D90747  
-----