
Midmark M11 - v1.0.0
Total Cycles: 2249

Name: RITTER 1
Sterilizer ID: V2178198
08/25/2022 01:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.3 F 0.0
2:00 199.2 F 0.7
4:00 213.5 F 3.6
6:00 228.3 F 8.9
8:00 238.1 F 9.9
10:00 240.8 F 11.8
12:00 239.6 F 10.2
14:00 237.9 F 9.1
16:00 252.9 F 16.6
18:00 265.3 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 272.7 F 28.9
2:00 271.8 F 28.0
3:00 272.8 F 29.0
4:00 272.0 F 28.2

Min 271.5 F 27.9
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 237.9 F 0.0
10:00 234.8 F 0.0
15:00 228.3 F 0.0
20:00 227.1 F 0.0
25:00 229.5 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 1:51
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:35
CYCLE COMPLETE

08/25/2022 02:04 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____