

-----  
Brampton, ON  
Canada  
-----

STATIM 5000 S501R713  
SN 511019D00004  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 001491  
11:55 10/12/2021  
WRAPPED  
134 C for 6 min.  
CYCLE START 0:00  
CONDITIONING 4:11  
PRESSURIZING 13:13  
STERILIZING 14:29  
136.0 C 220kPa 14:29  
136.4 C 220kPa 15:00  
136.4 C 222kPa 15:30  
136.4 C 223kPa 16:00  
136.5 C 220kPa 16:30  
136.3 C 222kPa 17:00  
135.8 C 221kPa 17:30  
136.2 C 220kPa 18:00  
136.4 C 222kPa 18:30  
136.3 C 223kPa 19:00  
136.1 C 222kPa 19:30  
136.4 C 223kPa 20:00  
136.4 C 222kPa 20:29  
VENTING 20:29  
AIR DRYING 21:26  
CYCLE COMPLETE 51:26  
-----

Digital Signature #  
E10F22BCE61684BC  
-----