
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005346
9:29 11/04/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:07
PRESSURIZING 2:10
STERILIZING 3:06
136.5 C 222kPa 3:06
136.1 C 224kPa 3:36
136.2 C 224kPa 4:06
136.4 C 222kPa 4:36
136.4 C 225kPa 5:06
136.4 C 225kPa 5:36
136.4 C 225kPa 6:06
136.4 C 222kPa 6:36
136.4 C 222kPa 6:36
VENTING 6:36
AIR DRYING 7:27
CYCLE COMPLETE 8:32

Digital Signature #
F26F652FBB3BA0D8
