
Midmark M11 - v1.0.5
Total Cycles: 1370

Name: RITTER2
Sterilizer ID: V2313228
08/04/2022 01:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 140.8 F 0.0
2:00 142.1 F 0.0
4:00 155.1 F 0.3
6:00 184.7 F 3.4
8:00 206.4 F 7.6
10:00 225.4 F 12.3
12:00 226.8 F 8.0
14:00 237.8 F 13.1
16:00 238.5 F 10.9
18:00 236.8 F 9.4
20:00 250.3 F 16.0
22:00 261.6 F 22.4
24:00 271.8 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.0
1:00 272.6 F 29.8
2:00 272.2 F 29.2
3:00 271.5 F 29.1
4:00 272.7 F 29.8

Min 271.4 F 28.8
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 206.0 F 0.0
10:00 213.4 F 0.0
15:00 215.2 F 0.0
20:00 215.0 F 0.0
25:00 217.1 F 0.0
30:00 217.3 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 24:38
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 01:02:57
CYCLE COMPLETE

08/04/2022 02:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____