
Midmark M11 - v1.0.0
Total Cycles: 4842

Name: RITTER 1
Sterilizer ID: V2178198
08/22/2024 09:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.6 F 0.0
2:00 158.8 F 0.5
4:00 159.1 F 0.9
6:00 186.5 F 2.4
8:00 212.1 F 6.3
10:00 220.2 F 10.4
12:00 229.4 F 7.5
14:00 233.9 F 10.0
16:00 237.8 F 9.0
18:00 244.2 F 12.8
20:00 242.3 F 11.3
22:00 254.8 F 17.6
24:00 265.4 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.4 F 28.9
2:00 271.8 F 28.3
3:00 271.6 F 28.6
4:00 272.4 F 28.7

Min 271.2 F 28.0
Max 272.5 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 201.2 F 0.0
10:00 200.5 F 0.0
15:00 219.9 F 0.0
20:00 237.6 F 0.0
25:00 229.4 F 0.0
30:00 231.7 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 25:53
STERILIZING: 4:00
VENTING: 3:39
DRYING: 30:00
TOTAL CYCLE: 01:04:19
CYCLE COMPLETE

08/22/2024 10:54 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____