
Midmark M11 - v1.0.0
Total Cycles: 901

Name: RITTER 1
Sterilizer ID: V2178198
06/15/2021 02:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 164.5 F 0.0
2:00 165.9 F 2.1
4:00 171.5 F 5.0
6:00 191.8 F 9.1
8:00 212.5 F 9.8
10:00 222.5 F 7.2
12:00 236.5 F 11.9
14:00 234.8 F 8.7
16:00 241.2 F 10.7
18:00 245.4 F 13.0
20:00 257.4 F 19.4
22:00 268.2 F 26.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.0
2:00 271.4 F 28.4
3:00 272.4 F 29.1
4:00 271.7 F 28.5

Min 271.3 F 28.3
Max 272.5 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.5 F 0.0
5:00 210.6 F 0.0
10:00 224.5 F 0.0
15:00 236.5 F 0.0
20:00 235.3 F 0.0
25:00 230.3 F 0.0
30:00 224.8 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 23:18
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:01:33
CYCLE COMPLETE

06/15/2021 03:26 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____