
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000532
14:47 17/06/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:48
STERILIZING 5:07
136.0 C 229kPa 5:07
136.2 C 231kPa 5:37
136.4 C 233kPa 6:07
136.4 C 229kPa 6:37
136.2 C 232kPa 7:07
136.3 C 231kPa 7:37
136.4 C 233kPa 8:07
136.3 C 231kPa 8:37
136.3 C 231kPa 8:37
VENTING 8:38
AIR DRYING 9:33
CYCLE COMPLETE 24:33

Digital Signature #
EE4A700F935CF1FB
