
Midmark M11 - v1.0.5
Total Cycles: 4377

Name: RITTER2
Sterilizer ID: V2313228
02/25/2025 10:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.2 F 0.0
2:00 71.7 F 0.2
4:00 74.2 F 1.0
6:00 89.8 F 2.9
8:00 131.8 F 5.4
10:00 156.8 F 8.5
12:00 174.1 F 11.9
14:00 201.1 F 3.1
16:00 212.4 F 6.4
18:00 222.4 F 9.8
20:00 231.8 F 12.5
22:00 232.7 F 9.4
24:00 241.1 F 11.5
26:00 242.2 F 12.2
28:00 252.9 F 17.8
30:00 262.7 F 23.6
32:00 272.2 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.5
1:00 272.4 F 30.3
2:00 272.3 F 29.9
3:00 271.1 F 29.4
4:00 271.5 F 30.1

Min 271.0 F 29.4
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.8 F 0.0
5:00 214.3 F 0.0
10:00 223.4 F 0.0
15:00 175.2 F 0.0
20:00 167.2 F 0.0
25:00 166.4 F 0.0
30:00 164.7 F 0.0

Drying Complete

FILLING: 1:03
HEATING: 32:34
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:11:28
CYCLE COMPLETE

02/25/2025 11:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____