
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000323
8:47 11/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:50
PRESSURIZING 2:59
STERILIZING 4:30
136.4 C 231kPa 4:30
136.4 C 232kPa 5:00
136.2 C 230kPa 5:30
136.3 C 230kPa 6:00
136.0 C 230kPa 6:30
136.4 C 233kPa 7:00
136.4 C 232kPa 7:30
136.4 C 233kPa 8:00
136.4 C 233kPa 8:00
VENTING 8:01
AIR DRYING 8:53
CYCLE COMPLETE 23:53

Digital Signature #
E87DA27421AF3253
