
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 000066
13:28 03/07/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:01
CYCLE ABORTED 3:07
WATER LEVEL LOW
NOT STERILE
END TIME 3:07

Digital Signature #
678D7D4DBE8EE212
