
Midmark M11 - v1.0.0
Total Cycles: 3412

Name: RITTER 1
Sterilizer ID: V2178198
06/20/2023 10:11 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 152.9 F 0.0
2:00 156.3 F 0.0
4:00 172.4 F 2.2
6:00 196.5 F 6.3
8:00 217.9 F 11.4
10:00 224.4 F 6.5
12:00 238.0 F 11.9
14:00 236.7 F 9.4
16:00 238.5 F 9.1
18:00 248.9 F 14.5
20:00 260.7 F 21.1
22:00 271.5 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.2 F 28.6
2:00 272.1 F 28.8
3:00 271.9 F 28.3
4:00 272.4 F 29.0

Min 271.4 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.1 F 0.0
5:00 213.6 F 0.0
10:00 216.9 F 0.0
15:00 220.0 F 0.0
20:00 217.8 F 0.0
25:00 217.1 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 0:48
HEATING: 22:35
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

06/20/2023 11:11 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____