
Midmark M11 - v1.0.0
Total Cycles: 2240

Name: RITTER 1
Sterilizer ID: V2178198
08/24/2022 01:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 137.4 F 0.0
2:00 137.9 F 0.1
4:00 156.5 F 0.5
6:00 184.0 F 2.5
8:00 205.1 F 6.9
10:00 224.9 F 11.5
12:00 227.3 F 6.9
14:00 239.9 F 12.1
16:00 237.2 F 9.6
18:00 238.2 F 8.8
20:00 249.1 F 14.7
22:00 261.3 F 21.6
24:00 272.4 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.5 F 28.3
3:00 272.5 F 28.8
4:00 271.5 F 28.2

Min 271.4 F 28.1
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 200.4 F 0.0
5:00 230.2 F 0.0
10:00 201.6 F 0.0
15:00 220.8 F 0.0
20:00 235.2 F 0.0
25:00 240.1 F 0.0
30:00 230.5 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 24:28
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

08/24/2022 02:42 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____