
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000837
12:43 05/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:38
PRESSURIZING 3:43
STERILIZING 5:16
136.4 C 230kPa 5:16
136.1 C 229kPa 5:46
136.4 C 231kPa 6:16
136.2 C 232kPa 6:46
136.2 C 230kPa 7:16
136.3 C 231kPa 7:46
136.4 C 232kPa 8:16
136.3 C 231kPa 8:46
136.3 C 231kPa 8:46
VENTING 8:46
AIR DRYING 9:40
CYCLE COMPLETE 24:40

Digital Signature #
B3330A41BD6F470A
