
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000396
13:05 25/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:25
PRESSURIZING 3:26
STERILIZING 4:48
136.4 C 233kPa 4:48
136.4 C 233kPa 5:18
136.3 C 233kPa 5:48
136.3 C 233kPa 6:18
136.3 C 231kPa 6:48
136.4 C 233kPa 7:18
136.4 C 233kPa 7:48
136.2 C 232kPa 8:18
136.2 C 232kPa 8:18
VENTING 8:18
AIR DRYING 9:14
CYCLE COMPLETE 9:15

Digital Signature #
6D2A8670F1907BC1
