

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 001956  
15:56 14/08/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:49  
PRESSURIZING 1:54  
STERILIZING 2:55  
136.4 C 226kPa 2:55  
136.4 C 225kPa 3:25  
136.4 C 225kPa 3:55  
136.2 C 223kPa 4:25  
136.3 C 225kPa 4:55  
136.3 C 224kPa 5:25  
136.4 C 225kPa 5:55  
136.4 C 225kPa 6:25  
136.4 C 225kPa 6:25  
VENTING 6:25  
AIR DRYING 7:15  
CYCLE COMPLETE 7:19  
-----

Digital Signature #  
EB9B48138D597AC1  
-----