
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000843
9:26 06/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:47
PRESSURIZING 3:51
STERILIZING 5:23
136.0 C 228kPa 5:23
136.2 C 231kPa 5:53
136.2 C 231kPa 6:23
136.3 C 230kPa 6:53
136.4 C 233kPa 7:23
136.3 C 231kPa 7:53
136.0 C 229kPa 8:23
136.4 C 231kPa 8:53
136.4 C 231kPa 8:53
VENTING 8:53
AIR DRYING 9:50
CYCLE COMPLETE 24:50

Digital Signature #
26F363608F2E3280
