
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000359
9:45 20/08/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:16
STERILIZING 4:19
136.4 C 224kPa 4:19
136.2 C 223kPa 4:49
136.4 C 223kPa 5:19
136.3 C 223kPa 5:49
136.2 C 222kPa 6:19
136.1 C 220kPa 6:49
136.2 C 220kPa 7:19
136.4 C 221kPa 7:49
136.4 C 221kPa 7:49
VENTING 7:49
AIR DRYING 8:46
CYCLE COMPLETE 10:35

Digital Signature #
B71B25BB869477EC
