
Midmark M11 - v1.0.0
Total Cycles: 4607

Name: RITTER 1
Sterilizer ID: V2178198
06/04/2024 12:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 146.9 F 0.0
2:00 150.1 F 2.6
4:00 162.7 F 5.8
6:00 197.1 F 9.8
8:00 219.9 F 5.5
10:00 225.9 F 7.4
12:00 235.2 F 12.1
14:00 233.7 F 8.3
16:00 242.9 F 12.1
18:00 243.7 F 12.2
20:00 255.6 F 18.3
22:00 266.7 F 25.1
24:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.3 F 28.7
2:00 272.4 F 29.1
3:00 272.1 F 29.1
4:00 272.1 F 28.7

Min 271.5 F 28.6
Max 273.0 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.2 F 0.0
5:00 234.1 F 0.0
10:00 252.9 F 0.0
15:00 231.2 F 0.0
20:00 209.8 F 0.0
25:00 208.4 F 0.0
30:00 215.0 F 0.0
Drying Complete

FILLING: 1:03
HEATING: 24:10
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:02:44
CYCLE COMPLETE

06/04/2024 01:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____