
Midmark M11 - v1.0.5
Total Cycles: 4393

Name: RITTER2
Sterilizer ID: V2313228
02/28/2025 04:12 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 167.7 F 0.0
2:00 170.0 F 3.3
4:00 187.8 F 8.0
6:00 211.9 F 13.2
8:00 223.4 F 7.2
10:00 235.8 F 12.3
12:00 238.3 F 10.4
14:00 236.4 F 9.2
16:00 250.2 F 15.8
18:00 261.6 F 22.4
20:00 272.0 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 272.6 F 29.6
1:00 272.7 F 29.8
2:00 272.4 F 30.1
3:00 272.0 F 30.1
4:00 271.9 F 29.5

Min 271.7 F 29.5
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 200.6 F 0.0
10:00 191.4 F 0.0
15:00 183.6 F 0.0
20:00 178.2 F 0.0
25:00 178.7 F 0.0
30:00 179.6 F 0.0
Drying Complete

FILLING: 1:46
HEATING: 21:09
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:00:45
CYCLE COMPLETE

02/28/2025 05:12 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____