
Midmark M11 - v1.0.0
Total Cycles: 5064

Name: RITTER 1
Sterilizer ID: V2178198
11/06/2024 04:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 86.0 F 0.0
2:00 87.9 F 0.0
4:00 109.1 F 0.2
6:00 140.7 F 0.7
8:00 171.0 F 2.2
10:00 192.8 F 5.5
12:00 208.7 F 9.5
14:00 219.7 F 13.5
16:00 224.6 F 7.1
18:00 235.7 F 11.6
20:00 233.1 F 8.1
22:00 244.3 F 12.4
24:00 243.9 F 12.2
26:00 255.9 F 18.4
28:00 266.7 F 24.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.6 F 28.9
2:00 271.5 F 28.1
3:00 272.4 F 29.0
4:00 272.1 F 28.5

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.2 F 0.0
5:00 191.9 F 0.0
10:00 208.1 F 0.0
15:00 229.4 F 0.0
20:00 217.5 F 0.0
25:00 212.9 F 0.0
30:00 216.4 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:37
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:07:54
CYCLE COMPLETE

11/06/2024 05:34 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____