
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 001918
9:42 07/08/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:00
STERILIZING 3:01
136.0 C 220kPa 3:01
136.2 C 224kPa 3:31
136.1 C 221kPa 4:01
136.4 C 226kPa 4:31
136.3 C 223kPa 5:01
136.2 C 224kPa 5:31
136.4 C 225kPa 6:01
136.1 C 222kPa 6:31
136.1 C 222kPa 6:31
VENTING 6:31
AIR DRYING 7:22
CYCLE COMPLETE 37:22

Digital Signature #
2BC0EE0B6BCF4CC5
