
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002939
10:24 12/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:13
PRESSURIZING 3:16
STERILIZING 4:25
136.4 C 222kPa 4:25
136.4 C 221kPa 4:55
136.2 C 224kPa 5:25
136.3 C 224kPa 5:55
136.2 C 222kPa 6:25
136.4 C 221kPa 6:55
136.3 C 223kPa 7:25
136.3 C 224kPa 7:55
136.3 C 224kPa 7:55
VENTING 7:55
AIR DRYING 8:52
CYCLE COMPLETE 9:05

Digital Signature #
CECB72ADCC002802
