
Midmark M11 - v1.0.0
Total Cycles: 662

Name: MIDMARK
Sterilizer ID: V2178198
02/10/2021 01:15 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 159.9 F 0.0
2:00 157.5 F 2.7
4:00 169.3 F 6.6
6:00 193.8 F 12.1
8:00 224.0 F 7.8
10:00 238.9 F 10.2
12:00 243.0 F 11.5
14:00 240.2 F 10.1
16:00 254.7 F 17.4
18:00 267.3 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 27.9
1:00 272.9 F 28.7
2:00 272.2 F 28.5
3:00 271.6 F 28.3
4:00 271.6 F 28.4

Min 271.0 F 27.8
Max 273.0 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.7 F 0.0
5:00 235.1 F 0.0
10:00 225.8 F 0.0
15:00 229.2 F 0.0
20:00 212.9 F 0.0
Dry Cycle Aborted

FILLING: 1:08
HEATING: 19:20
STERILIZING: 4:00
VENTING: 2:56
DRYING: 20:27
TOTAL CYCLE: 00:47:51

02/10/2021 02:02 PM
ACCEPTED: X
REJECTED:
Operator: U01