
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000998
12:44 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:35
STERILIZING 3:45
136.4 C 232kPa 3:45
136.4 C 233kPa 4:15
136.3 C 233kPa 4:45
136.2 C 231kPa 5:15
136.3 C 233kPa 5:45
136.3 C 231kPa 6:15
136.4 C 232kPa 6:45
136.3 C 233kPa 7:15
136.3 C 233kPa 7:15
VENTING 7:16
AIR DRYING 8:07
CYCLE COMPLETE 23:07

Digital Signature #
82A1F6EE473B020E
