
Midmark M11 - v1.0.0
Total Cycles: 4377

Name: RITTER 1
Sterilizer ID: V2178198
03/12/2024 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 121.1 F 0.0
2:00 123.4 F 1.3
4:00 141.0 F 3.8
6:00 174.8 F 7.0
8:00 202.4 F 11.3
10:00 216.5 F 3.1
12:00 220.9 F 7.1
14:00 232.1 F 11.2
16:00 229.2 F 6.8
18:00 242.6 F 12.3
20:00 240.2 F 10.4
22:00 252.2 F 16.2
24:00 262.6 F 22.3
26:00 272.5 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 28.7
2:00 272.5 F 29.1
3:00 271.5 F 28.4
4:00 272.5 F 29.0

Min 271.3 F 28.4
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 235.3 F 0.0
10:00 229.6 F 0.0
15:00 229.9 F 0.0
20:00 222.0 F 0.0
25:00 217.8 F 0.0
30:00 223.5 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 27:09
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:05:54
CYCLE COMPLETE

03/12/2024 12:56 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____