
Midmark M11 - v1.0.0
Total Cycles: 2173

Name: RITTER 1
Sterilizer ID: V2178198
07/25/2022 05:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.2 F 0.0
2:00 191.5 F 4.4
4:00 192.9 F 9.1
6:00 219.6 F 6.7
8:00 228.8 F 9.4
10:00 235.8 F 9.2
12:00 242.1 F 12.4
14:00 241.1 F 11.0
16:00 254.3 F 17.7
18:00 265.6 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.4
3:00 272.5 F 29.3
4:00 272.1 F 28.8

Min 271.4 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.7 F 0.0
5:00 236.3 F 0.0
10:00 230.5 F 0.0
15:00 226.6 F 0.0
20:00 226.2 F 0.0
25:00 228.6 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 19:45
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:37
CYCLE COMPLETE

07/25/2022 06:11 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____