
Midmark M11 - v1.0.0
Total Cycles: 4517

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2024 12:06 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 147.9 F 0.0
2:00 146.8 F 2.4
4:00 169.2 F 5.7
6:00 190.6 F 9.4
8:00 206.1 F 13.5
10:00 219.7 F 5.7
12:00 231.6 F 9.5
14:00 238.5 F 11.7
16:00 239.0 F 10.6
18:00 235.9 F 8.4
20:00 249.3 F 14.8
22:00 259.9 F 20.7
24:00 269.7 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.4 F 29.1
2:00 272.3 F 28.7
3:00 271.2 F 28.2
4:00 271.5 F 28.7

Min 271.1 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.7 F 0.0
5:00 234.4 F 0.0
10:00 216.3 F 0.0
15:00 206.9 F 0.0
20:00 216.1 F 0.0
25:00 217.1 F 0.0
30:00 205.1 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 25:05
STERILIZING: 4:00
VENTING: 3:50
DRYING: 30:00
TOTAL CYCLE: 01:04:05
CYCLE COMPLETE

05/02/2024 01:10 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____