
Midmark M11 - v1.0.5
Total Cycles: 2614

Name: RITTER2
Sterilizer ID: V2313228
07/06/2023 07:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.1 F 0.0
2:00 71.5 F 0.0
4:00 77.5 F 0.0
6:00 97.3 F 0.0
8:00 151.1 F 0.1
10:00 192.8 F 0.5
12:00 202.3 F 3.4
14:00 208.4 F 6.7
16:00 220.8 F 10.1
18:00 233.4 F 10.5
20:00 232.5 F 9.4
22:00 242.2 F 11.2
24:00 242.0 F 11.8
26:00 239.3 F 10.3
28:00 251.7 F 16.4
30:00 262.0 F 22.5
32:00 271.6 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.6 F 29.4
2:00 271.5 F 28.8
3:00 272.7 F 29.5
4:00 271.7 F 28.7

Min 271.5 F 28.6
Max 272.8 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.0 F 0.0
5:00 196.9 F 0.0
10:00 198.7 F 0.0
15:00 217.9 F 0.0
20:00 222.0 F 0.0
25:00 222.4 F 0.0
30:00 222.4 F 0.0

Drying Complete

FILLING: 0:38
HEATING: 32:37
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:10:25
CYCLE COMPLETE

07/06/2023 08:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____