
Midmark M11 - v1.0.0
Total Cycles: 2386

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2022 10:35 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 218.4 F 0.0
2:00 201.7 F 3.9
4:00 205.3 F 8.7
6:00 228.4 F 8.4
8:00 234.6 F 9.9
10:00 237.7 F 8.0
12:00 244.9 F 12.6
14:00 245.2 F 12.7
16:00 258.4 F 19.8
18:00 270.0 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.3
1:00 271.9 F 28.4
2:00 272.5 F 28.8
3:00 271.7 F 28.6
4:00 271.9 F 28.4

Min 271.5 F 28.1
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 220.2 F 0.0
5:00 237.6 F 0.0
10:00 227.4 F 0.0
15:00 226.3 F 0.0
20:00 226.6 F 0.0
25:00 230.1 F 0.0
30:00 229.2 F 0.0

Drying Complete

FILLING: 2:02
HEATING: 18:49
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:57:58
CYCLE COMPLETE

10/03/2022 11:32 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____