
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007961
12:07 10/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:32
STERILIZING 3:28
136.5 C 222kPa 3:28
136.4 C 224kPa 3:58
135.9 C 221kPa 4:28
136.4 C 220kPa 4:58
136.3 C 222kPa 5:28
136.1 C 219kPa 5:58
136.4 C 224kPa 6:28
135.9 C 221kPa 6:58
135.9 C 221kPa 6:58
VENTING 6:58
AIR DRYING 7:46
CYCLE COMPLETE 33:42

Digital Signature #
DE7A04AA602D8BEF
