

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3064

Name: RITTER 1  
Sterilizer ID: V2178198  
03/31/2023 09:32 AM

BEGIN UNWRAPPED CYCLE  
Temp: 270 Degrees F  
Time: 3 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 215.3 F 0.0  
2:00 211.8 F 5.6  
4:00 227.8 F 12.0  
6:00 236.1 F 9.9  
8:00 236.1 F 8.7  
10:00 237.3 F 8.5  
12:00 253.7 F 16.8

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 252F  
CHAMBER PRESS 15.3PSI  
\*\* ITEMS NOT STERILE \*\*

03/31/2023 09:49 AM