

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2757

Name: RITTER2  
Sterilizer ID: V2313228  
08/14/2023 09:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 135.7 F 0.0  
2:00 137.1 F 0.0  
4:00 136.8 F 0.0  
6:00 175.1 F 0.0  
8:00 193.6 F 0.0  
10:00 208.3 F 1.2  
12:00 218.5 F 4.7  
14:00 228.1 F 8.3  
16:00 237.5 F 12.3  
18:00 236.1 F 9.4  
20:00 241.1 F 10.9  
22:00 245.8 F 13.4  
24:00 244.8 F 12.7  
26:00 256.0 F 18.6  
28:00 266.1 F 24.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.5  
1:00 272.6 F 29.0  
2:00 272.1 F 29.0  
3:00 272.2 F 28.7  
4:00 272.5 F 29.3

Min 271.6 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.4 F 0.0  
5:00 227.3 F 0.0  
10:00 225.8 F 0.0  
15:00 220.4 F 0.0  
20:00 213.8 F 0.0  
25:00 212.6 F 0.0  
30:00 209.6 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 29:47  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:08:09  
CYCLE COMPLETE

08/14/2023 10:45 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_