

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005213  
11:39 13/03/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:52  
PRESSURIZING 1:57  
STERILIZING 2:45  
136.0 C 220kPa 2:45  
136.4 C 223kPa 3:15  
136.3 C 224kPa 3:45  
136.2 C 223kPa 4:15  
136.2 C 221kPa 4:45  
136.4 C 223kPa 5:15  
136.4 C 223kPa 5:45  
136.4 C 225kPa 6:15  
136.4 C 225kPa 6:15  
VENTING 6:15  
AIR DRYING 7:06  
CYCLE COMPLETE 10:51

-----  
Digital Signature #  
08519394C22BDC02  
-----